

TIANJIN CHEERMAN TECHNOLOGY CO., LTD

## CERTIFICATE OF SPECIFICATION

We claim:

DT8380 model is tested with OK accuracy to use

Model number:DT8380

Temperature:-50 — 380 °C



SPECIFICATION	
Temperature Range	-50~380°C
Accuracy Test Result	Good Accuracy
Test Times	3 times
Test Temperature Points	Test Results
0 to 100 °C	± 1°C
150 °C	153 °C
200 °C	203 °C
300 °C	305.50 °C
350 °C	355.50 °C

Approved By:

A handwritten signature in black ink, appearing to be "Tianjin".